



CERTIFICATE OF ACCREDITATION

ANSI-ASQ National Accreditation Board

500 Montgomery Street, Suite 625, Alexandria, VA 22314, 877-344-3044

This is to certify that

Cal-Chek Canada, Inc.
250 Governor's Road
Dundas ON L9H 3K3

has been assessed by ANAB
and meets the requirements of international standard

ISO/IEC 17025:2005

while demonstrating technical competence in the field of

CALIBRATION

Refer to the accompanying Scope of Accreditation for information regarding the types of calibrations to which this accreditation applies.

L1001-1
Certificate Number


ANAB Approval

Certificate Valid: 07/05/2018-08/11/2021
Version No. 003 Issued: 07/05/2018



This laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025:2005. This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (refer to joint ISO-ILAC-IAF Communiqué dated April 2017).



SCOPE OF ACCREDITATION TO ISO/IEC 17025:2005

Cal-Chek Canada, Inc.
 250 Governor's Road
 Dundas, ON L9H 3K3
 Kevin Newitt 905-628-4636

CALIBRATION

Valid to: **August 11, 2021**

Certificate Number: **L1001-1**

Length – Dimensional Metrology

Parameter / Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method and/or Equipment
Extrusion Plastometers (Melt Indexers, Melt Flow Indexers) Piston Foot Length Piston Foot Diameter PPDT Switch Calibration Go / No-Go Verification Bore Measurement Orifice Length	(0.2 to 0.3) in	1 700 µin	ASTM D1238 using: Caliper
	(0.3 to 0.4) in	120 µin	Micrometer
	(0.2 to 1.2) in	1 200 µin	Micrometer Head
	(0.082 to 0.083) in	110 µin	Micrometer
	(0.3 to 0.4) in	180 µin	Bore Gauge
	(0.3 to 0.4) in	120 µin	Micrometer
Extensometer Systems (Strain Instruments, Extensometers, Deflectometers)	(0.000 1 to 1) in	(18 + 59L) µin	ASTM E83 using: w/Cal-60
	(0.005 to 17) in	(94 + 27L) µin	Gauge blocks
Displacement Measuring Systems and Devices	(0.005 to 17) in	(18 + 59L) µin	ASTM E83 using: LVDT Calibrator
	(0.0001 to 1) in	(150 + 540L) µin	Dial Gauge
	(0.005 to 3) in	(1 100 + 180L) µin	Gauge blocks



Mass and Mass Related

Parameter / Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method and/or Equipment
Force Testing Machines Compression	(0.1 to 61) lbf (61 to 600 000) lbf	(0.001 1 + 0.000 3M) lbf 0.11% of Applied Load	ASTM E4, ASTM C39, CSA A23.2-9C using: Dead Weights Load Cells and Readout
Force Testing Machines Tension	(0.1 to 61) lbf (61 to 300 000) lbf	(0.001 1 + 0.000 3M) lbf 0.11% of Applied Load	ASTM E4, CSA A23.2-9C using: Dead Weights Load Cells and Readout
Brinell Hardness Tester	(500 to 3 000) kgf	4.1 kgf	ASTM E10 – Direct Verification using Brinell Proving Ring
Brinell Hardness Tester HBW @ 3 000 kgf	Low Medium High	2.5 HBW 5 HBW 8.7 HBW	ASTM E10 – Indirect Verification using Standardized Test Blocks
Rockwell Hardness Testers	HRA Low Medium High HRBW Low Medium High HRC Low Medium High HRFW Low Medium High HRRW 118	0.43 HRA 0.2 HRA 0.21 HRA 0.67 HRBW 0.65 HRBW 0.46 HRBW 0.38 HRC 0.33 HRC 0.32 HRC 0.62 HRFW 0.46 HRFW 0.47 HRFW 0.32 HRRW	ASTM E18 – Indirect Verification by Standardized Test Blocks

Mass and Mass Related

Parameter / Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method and/or Equipment
Rockwell Superficial Hardness Testers	HR15N		ASTM E18 – Indirect Verification by Standardized Test Blocks
	Low	0.44 HR15N	
	Medium	0.24 HR15N	
	High	0.22 HR15N	
	HR15TW		
	Low	0.36 HR15TW	
	Medium	0.37 HR15TW	
	High	0.32 HR15TW	
	HR30N		
	Low	0.41 HR30N	
	Medium	0.19 HR30N	
	High	0.30 HR30N	
	HR30TW		
	Low	0.52 HR30TW	
Medium	0.31 HR30TW		
High	0.33 HR30TW		
HR45N			
Low	0.54 HR45N		
Medium	0.58 HR45N		
High	0.29 HR45N		
HR45TW			
Low	0.70 HR45TW		
Medium	0.61 HR45TW		
High	0.46 HR45TW		
HR15YW			
90	0.98 HR15YW		
Extrusion Plastometers: Weights	(90 to 12 000) g	0.15% of Weight	ASTM D1238 using: Bench Scale

Thermodynamic

Parameter / Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method and/or Equipment
Extrusion Plastometers: Temperature Control Systems	(20 to 400) °C	0.08 °C	ASTM D1238

Time and Frequency

Parameter / Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method and/or Equipment
Extrusion Plastometers: Time Devices/Timers	(10 to 600) S	1.2 S	ASTM D1238

Calibration and Measurement Capability (CMC) is expressed in terms of the measurement parameter, measurement range, expanded uncertainty of measurement and reference standard, method, and/or equipment. The expanded uncertainty of measurement is expressed as the standard uncertainty of the measurement multiplied by a coverage factor of 2 ($k=2$), corresponding to a confidence level of approximately 95%.

Notes:

1. On-site calibration service is available for this parameter, since on-site conditions are typically more variable than those in the laboratory, larger measurement uncertainties are expected on-site than what is reported on the accredited scope.
2. M = force in lbf, L = length in inches.
3. This scope is formatted as part of a single document including Certificate of Accreditation No. L1001-1.



Vice President